

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N <u>130 289</u>	
Supervisor/Operator Performing the Verification Procedure:	
Name <u>James K. Kimura</u>	ID# <u>5677</u> Date <u>8/1/2011</u>
A Agency <u>Alaska State Troopers</u> Phone # <u>225-5118</u>	
Instrument Location <u>Ketchikan AST Post</u>	
B Alco S/N <u>78268</u> Target Value <u>.081</u> High Pressure <u>400 PSI</u>	
Alco Test Values <u>.083</u> <u>.082</u> 1st Alco 2nd Alco	
Signature <u>[Signature]</u>	
(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of Aug, 2011.

[Signature] (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130289

BMB
8/3/11

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

AUGUST 01, 2011

OPERATOR'S NAME:

KIMURA/JAMES/K

OPERATOR'S NUMBER: 5677

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 0000000

DEPT/AGENCY: KETO

CASE/REPORT: 11-000000

TEST TYPE: V

ALCO TARGET VALUE: .081

ALCO S/N: 78268

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 30.17 in

ALCO TARGET .081 00:37

BLANK TEST .000 00:38

INTERNAL STANDARD VERIFIED 00:38

ALCO TV 30.17 in .083 00:38

BLANK TEST .000 00:39

SUBJECT SAMPLE .000 00:39

BLANK TEST .000 00:40

ALCO TV 30.19 in .082 00:40

BLANK TEST .000 00:41

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

AUGUST 01, 2011

TIME 00:42

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 01/00/09

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 42c

BAROMETER: 30.17 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmno
pqrstuvwxyz{|}~"